

EU Declaration of Conformity

Expro Meters, Inc
50 Barnes Park North
Wallingford, Connecticut 06492 USA

do herewith declare that the **TPS Processing Module ("TPM")**

with model numbers **TPM-B-A-x-y-z**
(where x represents 1 or more alphanumeric characters
and y represents "1" or "2"
and z represents any alphanumeric string containing the letter "J")

is in conformity with the provisions of the following EU Directives when installed in accordance with the installation instructions provided with the TPM:

2014/35/EU Low Voltage Directive (LVD)

2014/30/EU Electromagnetic Compatibility (EMC) Directive
[TPM is categorized as Group 1, Class A ISM equipment.]

and that the standards referenced below have been applied:


EN 61010-1:2010 Safety Requirements for Electrical Equipment for Measurement,
Control, and Laboratory Use – Part 1: General Requirements

EN 55011:2009 Industrial, Scientific, and Medical (ISM) Radio-Frequency
+ A1:2010 Equipment – Electromagnetic Disturbance Characteristics – Limits
and Methods of Measurement (CISPR 11:2003)

EN 61326-1:2013 Electrical Equipment for Measurement, Control and Laboratory
Use – EMC Requirements – Part 1: General Requirements

EN 61326-2-3:2013 Electrical Equipment for Measurement, Control and Laboratory
Use – EMC Requirements – Part 2-3: ... Transducers ...

EN 60529:1991 Degrees of Protection Provided by Enclosure (IP Code)
+A1:2000

Name: Patrick Curry Title: General Manager
Signature:  Expro Meters, Inc
Date: 20 July 2015 50 Barnes Park North
Wallingford, CT 06492 USA

Year of CE Marking: 2011
